



## DICHIARAZIONE DI CONFORMITA'



Ai sensi del Regolamento UE 2017/745

L'azienda Demarta Virginio S.r.l. dichiara sotto la sua esclusiva responsabilità che i prodotti fabbricati ed immessi in commercio dalla stessa Demarta Virginio S.r.l.:

BASIC UDI-DI FODERE IN POLIURETANO: 8052141080591AJ9

Articolo	Codice	UDI-DI / EAN
Fodera in POLIURETANO 58x35x8 cm	05.91.58358	08052141083178
Fodera in POLIURETANO 190x80x10 cm	05.91.1908010	08052141088319
Fodera in POLIURETANO 190x80x13 cm	05.91.1908013	08052141083185
Fodera in POLIURETANO 190x80x14 cm	05.91.1908014	08052141083192
Fodera in POLIURETANO 190x80x15 cm	05.91.1908015	08052141083208
Fodera in POLIURETANO 190x85x10 cm	05.91.1908510	08052141083215
Fodera in POLIURETANO 190x85x14 cm	05.91.1908514	08052141083222
Fodera in POLIURETANO 195x85x14 cm	05.91.1958514	08052141083239
Fodera in POLIURETANO 195x120x14 cm	05.91.19512014	08052141083246
Fodera in POLIURETANO 198x88x14 cm	05.91.1988814	08052141083253

sono conformi alle disposizioni applicabili del **Regolamento 2017/745 sui Dispositivi Medici del 5 aprile 2017**.

La Demarta Virginio S.r.l. dichiara, inoltre, quanto segue:

- I dispositivi soddisfano i requisiti generali di sicurezza e prestazione così come richiesti dall'allegato I del regolamento 2017/745 come prescritto dall'allegato IV del suddetto regolamento.
- I dispositivi in tabella **NON SONO STRUMENTI DI MISURA**.
- I dispositivi in tabella **NON SONO DESTINATI AD INDAGINI CLINICHE**.
- I dispositivi in tabella vengono commercializzati in **confezione NON STERILE**.
- I dispositivi in tabella sono da considerarsi come appartenenti alla **classe I** in conformità a quanto stabilito dall'allegato VIII del suddetto regolamento.

Cerrione, 01/06/2023

DEMARTA VIRGINIO S.r.l.  
Dott. Virginio Belletti



## DECLARATION OF CONFORMITY



### Regulation 2017/745 on MEDICAL DEVICES of 5 April 2017

Demarta Virginio S.r.l. declares under its sole responsibility that the products made and traded by Demarta Virginio S.r.l. and belonging to the group of:

BASIC UDI-DI FOR POLYURETHANE COVER: 8052141080591AJ9

Article	Code	UDI-DI / EAN
POLYURETHANE Cover 58x35x8 cm	05.91.58358	08052141083178
POLYURETHANE Cover 190x80x10 cm	05.91.1908010	08052141088319
POLYURETHANE Cover 190x80x13 cm	05.91.1908013	08052141083185
POLYURETHANE Cover 190x80x14 cm	05.91.1908014	08052141083192
POLYURETHANE Cover 190x80x15 cm	05.91.1908015	08052141083208
POLYURETHANE Cover 190x85x10 cm	05.91.1908510	08052141083215
POLYURETHANE Cover 190x85x14 cm	05.91.1908514	08052141083222
POLYURETHANE Cover 195x85x14 cm	05.91.1958514	08052141083239
POLYURETHANE Cover 195x120x14 cm	05.91.19512014	08052141083246
POLYURETHANE Cover 198x88x14 cm	05.91.1988814	08052141083253

comply with the applicable provisions of the **2017/745 Regulation on Medical Devices of 5 April 2017**.

Demarta Virginio S.r.l. guarantees and declares under its sole responsibility what follows:

- The devices satisfy the requirements of general safety and performance requested by the Annex I of regulation 2017/745 as laid down by the Annex IV of the above mentioned regulation.
- The devices **ARE NOT MEASURING INSTRUMENTS**.
- The devices **ARE NOT MADE FOR CLINICAL TESTS**.
- The devices are packed in **NON-STERILE BOX**.
- The devices belong to **class I** in accordance with the provisions of Annex VIII of the above mentioned regulation.

Cerrione, 01/06/2023

DEMARTA VIRGINIO S.r.l.  
Dott. Virginio Belletti