

**Nome Prodotto**

Posizionatore Polso-Mano.

Codice	Misure (cm)
05.00.AP005	28 x 8.3 x 6

**Categoria**

Posizionatore in Gel per Sala Operatoria.

**Descrizione**

Il POSIZIONATORE POLSO-MANO permette di sostenere e posizionare l'articolazione del polso e la mano durante gli interventi chirurgici in queste parti del corpo.

**Materiali**

Il POSIZIONATORE POLSO-MANO è realizzato in gel amorfo di poliuretano con densità di 0,99 gr/cm<sup>3</sup> rivestito da un film di poliuretano.

**Manutenzione e Pulizia**

- I posizionatori in gel posso essere puliti facilmente con qualsiasi detergente e/o disinfettante per la sala operatoria.
- Evitare di immergere per lungo tempo i posizionatori in gel nel detergente e/o disinfettante.

**Conservazione e Smaltimento**

- Mantenere, se possibile, i posizionatori in una superficie in piano.
- Evitare di appoggiare sui posizionatori oggetti pesanti o taglienti.
- Mantenere la temperatura del gel tra i 50° e i -18°.
- Conservare al riparo da fonti di luce, umidità e calore.
- Smaltire il dispositivo rispettando le normative vigenti in materia di tutela ambientale e raccolta differenziata.

**Confezione**

1 Pezzo.

**Garanzia**

24 mesi per difetti di fabbricazione ad esclusione delle parti consumabili.

**Codice CND**

Y033399

**Codice RDM**

90736/R

**Classe Dispositivo Medico**

I conforme alla Direttiva MDR 2017/745



**Product Name**

Hand and Wrist Protection.

Code	Dimensions (cm)
05.00.AP005	28 x 8.3 x 6

**Category**

Gel Positioner for Operating Table.

**Description**

It can be used as hand and wrist support and protection during all types of surgery.

**Materials**It is made of an amorphous polyurethane gel with a density of 0.99 gr/cm<sup>3</sup>.**Maintenance and Cleaning**

- The gel positioners can be easily cleaned with any cleaning agent and / or disinfectant for the operating room.
- Avoid immersing the gel positioners in the detergent and / or disinfectant for a long time.

**Storage and Disposal**

- Keep the positioners, if possible, on a level surface.
- Avoid placing heavy or sharp objects on the positioners.
- Keep the gel temperature between 50 ° and -18 °.
- Keep away from sources of light, moisture and heat.
- Dispose of the device in compliance with current regulations on environmental protection and separate waste collection.

**Package**

1 Piece.

**Warranty**

24 months for manufacturing defects with the exception of consumable parts.

**CND Code**

Y033399

**RDM Code**

90736/R

**Medical Device Class**

I complies with Directive MDR 2017/745

