

**Nome Prodotto**

Cinture di Bloccaggio.

| Codice | Misure (cm) |
|------------|--------------|
| 05.00.FB01 | 33 x 7 x 1,5 |
| 05.00.FB02 | 36 x 7 x 1,5 |

Categoria

Posizionatore in Gel per Sala Operatoria.

Descrizione

Le CINTURE DI BLOCCAGGIO proteggono ed immobilizzano la parte del corpo in cui vengono posizionate (es. polso, gomito ecc.).

MaterialiLe CINTURE DI BLOCCAGGIO sono realizzate in gel amorfo di poliuretano con densità di 0,99 gr/cm³ rivestito da un film di poliuretano. La chiusura è in Velcro.**Manutenzione e Pulizia**

- I posizionatori in gel posso essere puliti facilmente con qualsiasi detergente e/o disinfettante per la sala operatoria.
- Evitare di immergere per lungo tempo i posizionatori in gel nel detergente e/o disinfettante.

Conservazione e Smaltimento

- Mantenere, se possibile, i posizionatori in una superficie in piano.
- Evitare di appoggiare sui posizionatori oggetti pesanti o taglienti.
- Mantenere la temperatura del gel tra i 50° e i -18°.
- Conservare al riparo da fonti di luce, umidità e calore.
- Smaltire il dispositivo rispettando le normative vigenti in materia di tutela ambientale e raccolta differenziata.

Confezione

2 Pezzi.

Garanzia

24 mesi per difetti di fabbricazione ad esclusione delle parti consumabili.

Codice CND

Y033399

Codice RDM

90736/R

Classe Dispositivo Medico

I conforme alla Direttiva MDR 2017/745



Product Name

Arm and Hand Stabilizer Strap.



| Code | Dimensions (cm) |
|------------|-----------------|
| 05.00.FB01 | 33 x 7 x 1.5 |
| 05.00.FB02 | 36 x 7 x 1.5 |

Category

Gel Positioner for Operating Table.

Description

They protect and clamp the body part in which they are placed (elbow, wrist etc..).

Materials

It is made of an amorphous polyurethane gel with a density of 0.99 gr/cm³. The closure is in Velcro.

Maintenance and Cleaning

- The gel positioners can be easily cleaned with any cleaning agent and / or disinfectant for the operating room.
- Avoid immersing the gel positioners in the detergent and / or disinfectant for a long time.

Storage and Disposal

- Keep the positioners, if possible, on a level surface.
- Avoid placing heavy or sharp objects on the positioners.
- Keep the gel temperature between 50 ° and -18 °.
- Keep away from sources of light, moisture and heat.
- Dispose of the device in compliance with current regulations on environmental protection and separate waste collection.

Package

2 Pieces.

Warranty

24 months for manufacturing defects with the exception of consumable parts.

CND Code

Y033399

RDM Code

90736/R

Medical Device Class

I complies with Directive MDR 2017/745

