

DICHIARAZIONE DI CONFORMITA'



Ai sensi del Regolamento UE 2017/745

L'azienda Demarta Virginio S.r.l. dichiara sotto la sua esclusiva responsabilità che i prodotti fabbricati ed immessi in commercio dalla stessa Demarta Virginio S.r.l.:

BASIC UDI-DI POLTRONE MECCANICHE: 8052141080111BG7

Articoli	Codice	UDI-DI / EAN
NOSTRESS Manuale Velluto Poliestere Marrone	01.11.Z1-VMM	08052141084502
NOSTRESS Manuale Pelle Color Antracite	01.11.Z1-PAM	08052141084519
NOSTRESS Manuale Pelle Color Lampone	01.11.Z1-PLM	08052141084526
NOSTRESS Manuale in Tessuto Poliestere Grigio	01.11.Z1-TGM	08052141084533
ESSENTIEL Manuale in Velluto Poliestere Marrone	01.11.Z3-VMM	08052141084625
ESSENTIEL Manuale in Pelle Color Cacao	01.11.Z3-PCM	08052141084632
ESSENTIEL Manuale in Poliestere Grigio	01.11.Z3-TGM	08052141084649
LIBERTY Manuale in Velluto Poliestere Marrone	01.11.Z4-VMM	08052141084687
LIBERTY Manuale in Pelle Color Cacao	01.11.Z4-PCM	08052141084694
LIBERTY Manuale in Pelle Color Lampone	01.11.Z4-PLM	08052141084700
LIBERTY Manuale in Poliestere Grigio	01.11.Z4-TGM	08052141084717
EVASION Manuale in Velluto Poliestere Marrone	01.11.Z6-VMM	8052141084809
EVASION Manuale in Pelle Color Cacao	01.11.Z6-PCM	8052141084816
EVASION Manuale in Pelle Color Lampone	01.11.Z6-PLM	8052141084823
EVASION Manuale in Poliestere Grigio	01.11.Z6-TGM	8052141084830

sono conformi alle disposizioni applicabili del **Regolamento 2017/745 sui Dispositivi Medici del 5 aprile 2017**.

La Demarta Virginio S.r.l. dichiara, inoltre, quanto segue:

- I dispositivi soddisfano i requisiti generali di sicurezza e prestazione così come richiesti dall'allegato I del regolamento 2017/745 come prescritto dall'allegato IV del suddetto regolamento.
- I dispositivi in tabella **NON SONO STRUMENTI DI MISURA**.
- I dispositivi in tabella **NON SONO DESTINATI AD INDAGINI CLINICHE**.
- I dispositivi in tabella vengono commercializzati in **confezione NON STERILE**.
- I dispositivi in tabella sono da considerarsi come appartenenti alla **classe I** in conformità a quanto stabilito dall'allegato VIII del suddetto regolamento.

Cerrione, 01/06/2023

DEMARTA VIRGINIO S.r.l.
Dott. Virginio Belletti





DECLARATION OF CONFORMITY



Regulation 2017/745 on MEDICAL DEVICES of 5 April 2017

Demarta Virginio S.r.l. declares under its sole responsibility that the products made and traded by Demarta Virginio S.r.l. and belonging to the group of:

BASIC UDI-DI FOR MECHANICAL LIFT CHAIRS: 8052141080111BG7

Articoli	Codice	UDI-DI / EAN
Manual NOSTRESS Brown Velvet	01.11.Z1-VMM	08052141084502
Manual NOSTRESS Anthracite Leather	01.11.Z1-PAM	08052141084519
Manual NOSTRESS Raspberry Leather	01.11.Z1-PLM	08052141084526
Manual NOSTRESS Gray Polyester	01.11.Z1-TGM	08052141084533
Manual ESSENTIEL Brown Velvet	01.11.Z3-VMM	08052141084625
Manual ESSENTIEL Cocoa-colored Leather	01.11.Z3-PCM	08052141084632
Manual ESSENTIEL Gray Polyester	01.11.Z3-TGM	08052141084649
Manual LIBERTY Brown Velvet	01.11.Z4-VMM	08052141084687
Manual LIBERTY Cocoa-colored Leather	01.11.Z4-PCM	08052141084694
Manual LIBERTY Raspberry Leather	01.11.Z4-PLM	08052141084700
Manual LIBERTY Gray Polyester	01.11.Z4-TGM	08052141084717
Manual EVASION Brown Velvet	01.11.Z6-VMM	8052141084809
Manual EVASION Cocoa-colored Leather	01.11.Z6-PCM	8052141084816
Manual EVASION Raspberry Leather	01.11.Z6-PLM	8052141084823
Manual EVASION Gray Polyester	01.11.Z6-TGM	8052141084830

comply with the applicable provisions of the **2017/745 Regulation on Medical Devices of 5 April 2017**.

Demarta Virginio S.r.l. guarantees and declares under its sole responsibility what follows:

- The devices satisfy the requirements of general safety and performance requested by the Annex I of regulation 2017/745 as laid down by the Annex IV of the above mentioned regulation.
- The devices **ARE NOT MEASURING INSTRUMENTS**.
- The devices **ARE NOT MADE FOR CLINICAL TESTS**.
- The devices are packed in **NON-STERILE BOX**.
- The devices belong to **class I** in accordance with the provisions of Annex VIII of the above mentioned regulation.

Cerrione, 01/06/2023

DEMARTA VIRGINIO S.r.l.
Dott. Virginio Belletti