

DICHIARAZIONE DI CONFORMITA'



Ai sensi del Regolamento UE 2017/745

L'azienda Demarta Virginio S.a.S. dichiara sotto la sua esclusiva responsabilità che i prodotti fabbricati ed immessi in commercio dalla stessa Demarta Virginio S.a.S.:

Articolo	Codice
Coppia di Supporti Ascellari	01.06.S
Coppia di Supporti Antibrachiali	01.06.A
Tavolino Antibrachiale	01.06.T
Sellino	01.06.510S
Ruote fisse • 7.5 cm (3")	01.06.B3
Ruote fisse • 10 cm (4")	01.06.B4
Ruote fisse • 12.7 cm (5")	01.06.B5
Ruote piroettanti • 7.5 cm (3")	01.06.A3
Ruote piroettanti • 10 cm (4")	01.06.A4
Ruote piroettanti • 12.7 cm (5")	01.06.A5
Puntali	01.06.C1
Puntali con ammortizzatore	01.06.C2

sono conformi alle disposizioni applicabili del **Regolamento 2017/745 sui Dispositivi Medici del 5 aprile 2017**.

La Demarta Virginio S.a.S. dichiara, inoltre, quanto segue:

- I dispositivi soddisfano i requisiti generali di sicurezza e prestazione così come richiesti dall'allegato I del regolamento 2017/745 come prescritto dall'allegato IV del suddetto regolamento.
- I dispositivi in tabella **NON SONO STRUMENTI DI MISURA**.
- I dispositivi in tabella **NON SONO DESTINATI AD INDAGINI CLINICHE**.
- I dispositivi in tabella vengono commercializzati in **confezione NON STERILE**.
- I dispositivi in tabella sono da considerarsi come appartenenti alla **classe I** in conformità a quanto stabilito dall'allegato VIII del suddetto regolamento.
- La Demarta Virginio S.a.S. mantiene e mette a disposizione delle Autorità Competenti, per almeno 10 anni dalla data di fabbricazione dell'ultimo lotto, la documentazione tecnica comprovante la conformità al regolamento 2017/745.

Pollone, 30/04/2021

DEMARTA VIRGINIO S.a.S.
Dott. Virginio Belletti





DECLARATION OF CONFORMITY



Regulation 2017/745 on MEDICAL DEVICES of 5 April 2017

Demarta Virginio S.a.S. declares under its sole responsibility that the products made and traded by Demarta Virginio S.a.S. and belonging to the group of:

Article	Code
Axillary supports	01.06.S
Forearm supports	01.06.A
Forearm support with table	01.06.T
Seat	01.06.510S
Fixed wheels • 7.5 cm (3")	01.06.B3
Fixed wheels • 10 cm (4")	01.06.B4
Fixed wheels • 12.7 cm (5")	01.06.B5
Swivel wheels • 7.5 cm (3")	01.06.A3
Swivel wheels • 10 cm (4")	01.06.A4
Swivel wheels • 12.7 cm (5")	01.06.A5
Tips	01.06.C1
Tips with shock absorber	01.06.C2

comply with the applicable provisions of the **2017/745 Regulation on Medical Devices of 5 April 2017**.

Demarta Virginio S.a.S. guarantees and declares under its sole responsibility what follows:

- The devices satisfy the requirements of general safety and performance requested by the Annex I of regulation 2017/745 as laid down by the Annex IV of the above mentioned regulation.
- The devices **ARE NOT MEASURING INSTRUMENTS**.
- The devices **ARE NOT MADE FOR CLINICAL TESTS**.
- The devices are packed in **NON-STERILE BOX**.
- The devices belong to **class I** in accordance with the provisions of Annex VIII of the above mentioned regulation.
- Demarta Virginio S.a.S. provides to the Competent Authorities the technical documentation to prove the conformity to the 2017/745 regulation, for at least 10 years from the last lot production.

Pollone, 30/04/2021

DEMARTA VIRGINIO S.a.S.
Dott. Virginio Belletti